

**WASHINGTON STATE PATROL
EVIDENTIARY SUBJECT TEST**
ALCOTEST 9510 SERIAL NUMBER ARKC-0074
SOFTWARE VERSION 8322798 0.9
CONFIGURATION VERSION 8322796 2.7
DATE OF LAST QAP: 07/30/2019

Analysis Date: **05/09/2020**
 Observation Period Began: **21:12**
 Citation/Case Number: **K20-003915**
 Operator Name: **JUSTIN/ TANNER/ D**
 Subject Name: **KELLER/ AUSTIN/ R**
 Subject Date of Birth: **02/24/1998**
 External Standard Lot: **999502**

Breath Analysis	Result g/210L	Time hh:mm	Volume liters	Blowtime seconds
Blank Test	0.000	21:32		
Internal Standard	VERIFIED	21:33		
Subject Sample 1			3.5	11.0
IR Result	0.117	21:34		
EC Result	0.116	21:34		
Blank Test	0.000	21:35		
External Standard IR	0.081	21:36		
External Standard EC	0.084	21:36		
Blank Test	0.000	21:37		
Subject Sample 2			3.8	12.1
IR Result	0.117	21:38		
EC Result	0.116	21:38		
Blank Test	0.000	21:39		

During this test, I followed all protocols set in place by the Washington State Toxicologist for the purposes of this test. At the time of this test I was certified to operate the Alcotest 9510 and possessed a valid permit issued by the State Toxicologist. I observed the subject during the entire observation period and during that time they did not eat, drink, smoke, vomit, or place any foreign substances in their mouth.

I certify (or declare) under penalty of perjury under the laws of the state of Washington that the statements on this document and information contained therein are true, correct, and accurate. (RCW 9A.72.085.)

Officer Signature Tanner D Date 5/9/2020
 Location Signed Silverdale, WA